



Apr 14th, 9:00 AM - 10:00 AM

# Scanning Electrochemical Microscopy Mapping of Neurotransmitter Release Sites from Model Neurons

Jennifer A. Schreiber  
*Illinois Wesleyan University*

Melinda Baur, Faculty Advisor  
*Illinois Wesleyan University*

Follow this and additional works at: <http://digitalcommons.iwu.edu/jwprc>

---

Schreiber, Jennifer A. and Baur, Faculty Advisor, Melinda, "Scanning Electrochemical Microscopy Mapping of Neurotransmitter Release Sites from Model Neurons" (2012). *John Wesley Powell Student Research Conference*. 17. <http://digitalcommons.iwu.edu/jwprc/2012/posters/17>

This Event is brought to you for free and open access by The Ames Library, the Andrew W. Mellon Center for Curricular and Faculty Development, the Office of the Provost and the Office of the President. It has been accepted for inclusion in Digital Commons @ IWU by the faculty at Illinois Wesleyan University. For more information, please contact [digitalcommons@iwu.edu](mailto:digitalcommons@iwu.edu).

©Copyright is owned by the author of this document.

Poster Presentation P33

**SCANNING ELECTROCHEMICAL MICROSCOPY MAPPING OF  
NEUROTRANSMITTER RELEASE SITES FROM MODEL NEURONS**

Jennifer A. Schreiber and Melinda Baur\*  
Chemistry Department, Illinois Wesleyan University

Scanning Electrochemical Microscopy (SECM) was used to image the topography of model neurons in the constant height mode while simultaneously measuring neurotransmitter concentration using amperometry in the collector mode. The probe of the SECM was a fabricated 5  $\mu\text{m}$  carbon-fiber electrode. The electrode was calibrated to determine the distance to the surface. Topographical imaging was accomplished by moving the electrode systematically across the surface. Multiple topographical images were obtained before and after stimulation of the model neurons with KCl. Neurotransmitters norepinephrine and dopamine were detected by setting the electrode to an oxidizing potential (+0.8 v). After stimulation, neurotransmitter release was detected amperometrically during topographical mapping. Download this file and use it as a template.