



Illinois Wesleyan University
Digital Commons @ IWU

John Wesley Powell Student Research
Conference

1994, 5th Annual JWP Conference

Apr 23rd, 9:00 AM - 4:00 PM

The Determination of Cocaine Metabolite in Urine Using High Performance Liquid Chromatography

Jason R. Babcock
Illinois Wesleyan University

David N. Bailey, Faculty Advisor
Illinois Wesleyan University

Follow this and additional works at: <https://digitalcommons.iwu.edu/jwprc>

Babcock, Jason R. and Bailey, Faculty Advisor, David N., "The Determination of Cocaine Metabolite in Urine Using High Performance Liquid Chromatography" (1994). *John Wesley Powell Student Research Conference*. 3.
<https://digitalcommons.iwu.edu/jwprc/1994/posters/3>

This Event is protected by copyright and/or related rights. It has been brought to you by Digital Commons @ IWU with permission from the rights-holder(s). You are free to use this material in any way that is permitted by the copyright and related rights legislation that applies to your use. For other uses you need to obtain permission from the rights-holder(s) directly, unless additional rights are indicated by a Creative Commons license in the record and/ or on the work itself. This material has been accepted for inclusion by faculty at Illinois Wesleyan University. For more information, please contact digitalcommons@iwu.edu.

©Copyright is owned by the author of this document.

**THE DETERMINATION OF COCAINE METABOLITE
IN URINE USING HIGH PERFORMANCE LIQUID
CHROMATOGRAPHY. Jason R. Babcock and
David N. Bailey*, Department of Chemistry
Illinois Wesleyan University Bloomington, Illinois 61701.**

The primary metabolite of cocaine found in the body after ingestion of the drug is benzoylecognine (BE). A problem arises in the detection of BE because it is highly water soluble and, therefore, difficult to extract from urine into a solvent. This problem makes cocaine one of the most difficult drugs to detect. Previous work has described a number of methods for the removal of BE from the urine followed by some form of analytical detection. The research in this paper focused on the use of an ion-pairing agent to aide in the removal of BE from aqueous solution. The ion-pairing complex used was a 0.05 M solution of hexakis(thiocyanato)iron(III), $\text{Fe}(\text{SCN})_6^{3-}$, with a maximum sensitivity achieved at $\text{pH}=2.7$. The urine was treated with the complex solution and immediately extracted with methylene chloride. The methylene chloride was then analyzed by High Performance Liquid Chromatography (HPLC) for the presence of the BE/complex ion pair. A detection limit well within the physiological concentration range was obtained.