



Illinois Wesleyan University
Digital Commons @ IWU

John Wesley Powell Student Research
Conference

2004, 15th Annual JWP Conference

Apr 17th, 9:00 AM - 10:00 AM

A Novel Method for Super-Resolution Optical Microscopy

Brett Parmenter
Illinois Wesleyan University

Gabriel C. Spalding, Faculty Advisor
Illinois Wesleyan University

Follow this and additional works at: <https://digitalcommons.iwu.edu/jwprc>

Parmenter, Brett and Spalding, Faculty Advisor, Gabriel C., "A Novel Method for Super-Resolution Optical Microscopy" (2004). *John Wesley Powell Student Research Conference*. 21.

<https://digitalcommons.iwu.edu/jwprc/2004/posters/21>

This Event is protected by copyright and/or related rights. It has been brought to you by Digital Commons @ IWU with permission from the rights-holder(s). You are free to use this material in any way that is permitted by the copyright and related rights legislation that applies to your use. For other uses you need to obtain permission from the rights-holder(s) directly, unless additional rights are indicated by a Creative Commons license in the record and/ or on the work itself. This material has been accepted for inclusion by faculty at Illinois Wesleyan University. For more information, please contact digitalcommons@iwu.edu.

©Copyright is owned by the author of this document.

Poster Presentation P37

A NOVEL METHOD FOR SUPER-RESOLUTION OPTICAL MICROSCOPY

Brett Parmenter and Gabriel C. Spalding*

Department of Physics, Illinois Wesleyan University
Department of Physics, Northwestern University

Even neglecting aberrations, conventional optical microscopes are limited (essentially by the uncertainty principle) to a resolution on the order of the wavelength of light used. The Near-Field Scanning Optical Microscope (NSOM) is able to overcome this fundamental limitation, by using evanescent fields. The most common NSOM design uses a fiber optic pulled down to a very fine tip with a sub-wavelength aperture at the end. The resolution of this type of microscope is limited only by the size of the aperture used. Unfortunately, as shown by Hans Bethe, the intensity of the light allowed through a sub-wavelength aperture falls off as the sixth power of the radius. Our alternative uses an optically trapped, resonant nanoparticle as a sub-wavelength, scanned probe. We expect to achieve super-resolution with improved signal-to-noise.