

Illinois Wesleyan University Digital Commons @ IWU

John Wesley Powell Student Research Conference

2011, 22nd Annual JWP Conference

Apr 9th, 2:35 PM - 3:35 PM

Particle Tracking and AOD Control

Nathaniel Wolanyk Illinois Wesleyan University

Carl Mueller Illinois Wesleyan University

Patrick Dahl Illinois Wesleyan University

Evan Baker Illinois Wesleyan University

Gabriel Spalding, Faculty Advisor Illinois Wesleyan University

Follow this and additional works at: https://digitalcommons.iwu.edu/jwprc



Part of the Physics Commons

Wolanyk, Nathaniel; Mueller, Carl; Dahl, Patrick; Baker, Evan; and Spalding, Faculty Advisor, Gabriel, "Particle Tracking and AOD Control" (2011). John Wesley Powell Student Research Conference. 24.

https://digitalcommons.iwu.edu/jwprc/2011/posters2/24

This Event is protected by copyright and/or related rights. It has been brought to you by Digital Commons @ IWU with permission from the rights-holder(s). You are free to use this material in any way that is permitted by the copyright and related rights legislation that applies to your use. For other uses you need to obtain permission from the rights-holder(s) directly, unless additional rights are indicated by a Creative Commons license in the record and/ or on the work itself. This material has been accepted for inclusion by faculty at Illinois Wesleyan University. For more information, please contact digitalcommons@iwu.edu.

©Copyright is owned by the author of this document.

Poster Presentation P52

PARTICLE TRACKING AND AOD CONTROL

Nathaniel Wolanyk, Carl Mueller, Patrick Dahl, Evan Baker and Gabriel Spalding*
Physics Department, Illinois Wesleyan University

My project is two-fold: to implement LabVIEW-based software control the digital I/O cards controlling two Acousto-Optic Deflectors (AOD) and to implement computer vision software for automated tracking of two microparticles and their proximity to each other. The AODs role in this project is to quickly switch the laser between two points and thus trap two microparticles with just one laser beam. The particles have no reason to move quickly because the only outside force on them is Brownian motion, due to the random thermal movement of water molecules. The goal of my AOD-control program was to quickly (under 1ms) switch from one trap to another and back again, while also equilibrating the power delivered to each trap. The basic control task has now been accomplished, by using a while loop to reiterate quickly, allowing rapid switching between the two different inputs, by using data acquisition (DAQ) sub-virtual instruments (subVIs) built into LabVIEW. As for particle tracking, we have begun use of a program developed by one of our collaborators that is quite advanced: bit-map frames captured during video microscopy are analyzed as an image sequence to identify and locate the particles in each frame. As a prequel, we have used additional software to convert raw video from the microscope to avi video, and finally to a bitmap image sequence. Also, I modified the particletracking program to display displacement, rather than speed. At this point, I am able to optically two microparticles, at controlled separations, and then, as the traps "blink" off, track their displacements, so as to build up Boltzmann statistics relevant to the particle-particle interactions.