



Apr 23rd, 9:00 AM - 4:00 PM

# The Determination of Cocaine Metabolite in Urine Using High Performance Liquid Chromatography

Jason R. Babcock

*Illinois Wesleyan University*

David N. Bailey, Faculty Advisor

*Illinois Wesleyan University*

Follow this and additional works at: <http://digitalcommons.iwu.edu/jwprc>

---

Babcock, Jason R. and Bailey, Faculty Advisor, David N., "The Determination of Cocaine Metabolite in Urine Using High Performance Liquid Chromatography" (1994). *John Wesley Powell Student Research Conference*. 3.  
<http://digitalcommons.iwu.edu/jwprc/1994/posters/3>

This Event is brought to you for free and open access by The Ames Library, the Andrew W. Mellon Center for Curricular and Faculty Development, the Office of the Provost and the Office of the President. It has been accepted for inclusion in Digital Commons @ IWU by the faculty at Illinois Wesleyan University. For more information, please contact [digitalcommons@iwu.edu](mailto:digitalcommons@iwu.edu).

©Copyright is owned by the author of this document.

**THE DETERMINATION OF COCAINE METABOLITE  
IN URINE USING HIGH PERFORMANCE LIQUID  
CHROMATOGRAPHY.** Jason R. Babcock and  
David N. Bailey\*, Department of Chemistry  
Illinois Wesleyan University Bloomington, Illinois 61701.

The primary metabolite of cocaine found in the body after ingestion of the drug is benzoylecognine (BE). A problem arises in the detection of BE because it is highly water soluble and, therefore, difficult to extract from urine into a solvent. This problem makes cocaine one of the most difficult drugs to detect. Previous work has described a number of methods for the removal of BE from the urine followed by some form of analytical detection. The research in this paper focused on the use of an ion-pairing agent to aide in the removal of BE from aqueous solution. The ion-pairing complex used was a 0.05 M solution of hexakis(thiocyanato)iron(III),  $\text{Fe}(\text{SCN})_6^{3-}$ , with a maximum sensitivity achieved at  $\text{pH}=2.7$ . The urine was treated with the complex solution and immediately extracted with methylene chloride. The methylene chloride was then analyzed by High Performance Liquid Chromatography (HPLC) for the presence of the BE/complex ion pair. A detection limit well within the physiological concentration range was obtained.